

Liberty® Cyclor Training Test

Please circle the best answer

- 1. When inserting the cassette, the plastic domes should be**
 - a. Facing inward
 - b. Facing outward
 - c. Up-side-down

- 2. The heater bag has the _____ color clamp**
 - a. Blue
 - b. Yellow
 - c. Red

- 3. It is necessary to mask and wash my hands before**
 - a. Connecting the solution bags to the tubing
 - b. Watching television
 - c. Turning on the cyclor

- 4. I should use _____ to clean the surfaces of the cyclor**
 - a. A damp cloth with a dilute bleach solution
 - b. Formula 409
 - c. Windex

- 5. When I travel, my Liberty cyclor can**
 - a. Be checked luggage on a plane
 - b. Be hand carried on the plane
 - c. Not go on a plane

- 6. When I experience a DRAIN ALARM, I will press STOP to mute the audible alarm and to display the HELP screen. After checking for possible causes of the alarm, I will**
 - a. Roll over and go back to sleep
 - b. Press stop again to interrupt the treatment for 10 minutes
 - c. Press OK to continue with the treatment

- 7. If the power fails during a treatment**
 - a. All lines will automatically be closed, the pump will stop, and the screen will be blank
 - b. An alarm will sound and a HELP screen will appear
 - c. Power failures cannot occur before starting your treatments

- 8. How do you bypass a phase of your treatment?**
 - a. You may not bypass any part of your treatment
 - b. Hold the OK key for four seconds and press "Bypass" on the screen
 - c. Bypassing only works when the machine is turned off



Liberty® Cycler Training Test

Please circle the best answer

- 9. It is important to check the solution bags before use. Check for**
- Solution, clarity, amount, leaks, expiration
 - Type of plastic, flexibility, amount
 - Clarity, color, cost
- 10. Under the 'hardware' tab of "My Settings" page, you can change**
- Letters
 - Alarm and key loudness, screen blanking and brightness
 - Percentage of power used
- 11. Under "Patient Data", I can enter my**
- Address
 - Blood pressure
 - Bank account number
- 12. If you are having problems with your Liberty cycler, you should call**
- The neighbors
 - Fresenius Technical Support
 - Your best friend
- 13. A bag will be attached to the line with the green clamp for the**
- Medication option
 - Last bag option
 - Heater bag option
- 14. Always check your prescription for**
- Correct information
 - Hand washing instructions
 - Number of unused lines
- 15. Your treatment will begin**
- Once you are lying down
 - Once you are connected and both catheter and blue patient line clamps are open and you press OK
 - Once the machine alarms twice
- 16. The stay-safe® organizer**
- Assists you in making the connection between your catheter and the blue patient line
 - Verifies that all of the clamps are open
 - Prevents solution from draining



Liberty® Cyclor Training Test

Please circle the best answer

- 17. When moving the cyclor**
- a. You can drag it by the cord
 - b. Push it with your foot
 - c. Use the handles on either side
- 18. One of the things that the OK button does on the machine is**
- a. It saves your settings
 - b. It goes back one screen
 - c. It puts the machine in "sleep mode"
- 19. If you are using an IQdrive, check that it**
- a. Has been recently cleaned
 - b. Is inserted properly into the slot on the back of your machine
 - c. Is free from obstructions
- 20. Be sure the power cord is**
- a. Connected to an extension cord
 - b. Plugged into a grounded, 3-prong electrical outlet
 - c. Plugged into the back of your television set
- 21. When can the cassette be inserted into the Liberty cyclor?**
- a. When the cyclor prompts you to "insert cassette"
 - b. Any time during a treatment
 - c. When the machine is turned off
- 22. For better visibility, when connecting the lines to the ports, the bags can be placed**
- a. On the ground
 - b. On the sink
 - c. On the heater tray
- 23. Where can I find a complete listing of alarms?**
- a. In the dictionary
 - b. In the Liberty Cyclor User Manual
 - c. In the Handi-Guide
- 24. Why do warnings usually occur during treatments?**
- a. To alert you about the conditions that need to be taken care of
 - b. To disturb your sleep
 - c. To initiate treatments



Liberty® Cyclor Training Test

Please circle the best answer

- 25. The range of height for the Liberty cyclor in your room is**
- a. Just above the floor
 - b. On the windowsill to the dresser
 - c. One foot higher or one foot lower than the bed height
- 26. For optimal performance, place the heater bag**
- a. In the suitcase
 - b. Between the lips on either side of the heater tray
 - c. Hanging over the edge of the heater tray
- 27. A BLUE status bar indicates**
- a. The door sensor is broken
 - b. Patient interaction is required
 - c. The next dwell is changing
- 28. If the temperature of the Delflex® bag exceeds 38°C, the Liberty cyclor will display**
- a. Please wait for the solution to cool down
 - b. A bypass has occurred
 - c. Press OK to start fill
- 29. "F" under My Records Details screen indicates**
- a. Fun for all
 - b. Fill up the supply bags
 - c. Power failed
- 30. The Liberty cyclor will not allow a large volume to be programmed**
- a. In a short amount of treatment time
 - b. As a time saver
 - c. If the cone is not broken



Fresenius Medical Care

Fresenius Medical Care North America
920 Winter Street
Waltham, MA 02451

PN# 101092-01 Rev. 00 06/2009

success@home™

